


NANOCOLOR[®] gesamt-Phosphat 5

total Phosphate / Phosphate total / Fosfato total



Test 0-81
REF 985081











Methode / Method / Méthode / Método

690 / 710 nm

0811 0.20–5.00 mg/L P


0812 0.5–15.0 mg/L PO₄³⁻

Min	20–80 %		Max
0.20	1.16	4.04	5.00 mg/L P
0.5	3.4	12.1	15.0 mg/L PO ₄ ³⁻



NANOCOLOR® ortho-Phosphat 5

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-81
REF 985081

Methode / Method / Méthode / Método

690 / 710 nm

0815 0.20–5.00 mg/L PO₄-P

0816 0.5–15.0 mg/L PO₄³⁻

Min	20–80 %		Max
0.20	1.16	4.04	5.00 mg/L PO ₄ -P
0.5	3.4	12.1	15.0 mg/L PO ₄ ³⁻

