


NANOCOLOR® Ammonium 3

Amonio



Test 0-03
REF 985003

Methode / Method / Méthode / Método

0031 0.04–2.30 mg/L $\text{NH}_4\text{-N}$

0032 0.05–3.00 mg/L NH_4^+

0033 0.05–3.00 mg/L NH_3

660 / 690 nm

Min	20–80 %		Max
0.04	0.49	1.85	2.30 mg/L $\text{NH}_4\text{-N}$
0.05	0.64	2.41	3.00 mg/L $\text{NH}_4^+ / \text{NH}_3$

